
Midmark M11 - v1.0.5
Total Cycles: 887

Name: RITTER2
Sterilizer ID: V2313228
03/16/2022 03:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.4 F 0.0
2:00 197.0 F 3.8
4:00 201.4 F 9.1
6:00 228.3 F 8.6
8:00 237.3 F 11.7
10:00 236.8 F 10.0
12:00 237.0 F 8.7
14:00 250.8 F 16.3
16:00 263.3 F 23.6
18:00 271.3 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.0 F 29.4
2:00 271.3 F 29.4
3:00 271.6 F 29.3
4:00 271.2 F 29.3

Min 271.0 F 28.9
Max 272.3 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.3 F 0.0
5:00 220.7 F 0.0
10:00 242.1 F 0.0
15:00 241.0 F 0.0
20:00 232.4 F 0.0
25:00 224.3 F 0.0
30:00 224.7 F 0.0
Drying Complete

FILLING: 2:17
HEATING: 18:01
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:57:24
CYCLE COMPLETE

03/16/2022 04:55 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____