
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 000379
8:26 01/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:48
PRESSURIZING 3:50
STERILIZING 4:57
136.0 C 220kPa 4:57
136.4 C 224kPa 5:27
136.2 C 221kPa 5:57
136.1 C 217kPa 6:27
136.3 C 222kPa 6:57
136.3 C 222kPa 7:27
136.1 C 221kPa 7:57
136.4 C 221kPa 8:27
136.4 C 221kPa 8:27
VENTING 8:27
AIR DRYING 9:22
CYCLE COMPLETE 9:26

Digital Signature #
36C9C1553B591E60
