
Midmark M11 - v1.0.0
Total Cycles: 1979

Name: RITTER 1
Sterilizer ID: V2178198
05/31/2022 10:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 135.9 F 0.0
2:00 138.3 F 1.7
4:00 156.5 F 5.1
6:00 192.5 F 9.2
8:00 207.3 F 7.9
10:00 222.0 F 7.9
12:00 232.8 F 12.5
14:00 233.1 F 8.4
16:00 243.1 F 11.9
18:00 242.5 F 11.7
20:00 254.7 F 17.9
22:00 265.5 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.3 F 29.0
2:00 271.1 F 28.4
3:00 272.3 F 29.1
4:00 271.4 F 28.4

Min 271.1 F 28.3
Max 272.4 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.7 F 0.0
5:00 224.3 F 0.0
10:00 223.6 F 0.0
15:00 217.4 F 0.0
20:00 230.9 F 0.0
25:00 230.0 F 0.0
30:00 228.8 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 23:48
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:01:60
CYCLE COMPLETE

05/31/2022 11:30 AM
ACCEPTED: X
REJECTED:
Operator: U02