
Midmark M11 - v1.0.0
Total Cycles: 2627

Name: RITTER 1
Sterilizer ID: V2178198
11/30/2022 10:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.1 F 0.0
2:00 189.0 F 3.9
4:00 195.2 F 8.8
6:00 215.5 F 8.6
8:00 227.7 F 9.0
10:00 235.0 F 9.4
12:00 242.6 F 12.6
14:00 241.8 F 11.5
16:00 255.6 F 18.5
18:00 267.5 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.5 F 29.3
2:00 271.7 F 28.6
3:00 272.3 F 29.4
4:00 272.3 F 29.1

Min 271.4 F 28.5
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.2 F 0.0
5:00 230.7 F 0.0
10:00 225.5 F 0.0
15:00 234.8 F 0.0
20:00 229.6 F 0.0
25:00 223.6 F 0.0
30:00 214.1 F 0.0
Drying Complete

FILLING: 1:28
HEATING: 19:22
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:02
CYCLE COMPLETE

11/30/2022 11:53 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____