
Midmark M11 - v1.0.5
Total Cycles: 987

Name: RITTER2
Sterilizer ID: V2313228
04/14/2022 06:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.8 F 0.0
2:00 77.6 F 0.0
4:00 100.9 F 0.0
6:00 124.3 F 0.3
8:00 155.3 F 3.8
10:00 179.3 F 7.7
12:00 199.1 F 12.1
14:00 214.5 F 5.3
16:00 226.3 F 9.6
18:00 233.8 F 10.2
20:00 239.5 F 11.6
22:00 238.4 F 10.2
24:00 252.0 F 17.0
26:00 263.6 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.2
1:00 272.5 F 30.0
2:00 271.6 F 29.0
3:00 272.6 F 29.8
4:00 271.6 F 29.2

Min 271.5 F 29.0
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.7 F 0.0
5:00 211.4 F 0.0
10:00 188.4 F 0.0
15:00 188.2 F 0.0
20:00 191.8 F 0.0
25:00 209.5 F 0.0
30:00 223.6 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 27:55
STERILIZING: 4:00
VENTING: 3:01
DRYING: 30:00
TOTAL CYCLE: 01:05:59
CYCLE COMPLETE

04/14/2022 07:58 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____