
Midmark M11 - v1.0.0
Total Cycles: 3413

Name: RITTER 1
Sterilizer ID: V2178198
06/20/2023 11:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.4 F 0.0
2:00 189.9 F 3.0
4:00 210.9 F 8.6
6:00 226.1 F 10.8
8:00 233.4 F 9.9
10:00 235.0 F 8.3
12:00 244.6 F 12.4
14:00 245.4 F 12.8
16:00 258.1 F 19.6
18:00 269.3 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.1
1:00 272.6 F 28.8
2:00 271.5 F 28.2
3:00 272.3 F 28.7
4:00 271.5 F 28.2

Min 271.4 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.9 F 0.0
5:00 217.7 F 0.0
10:00 225.3 F 0.0
15:00 231.1 F 0.0
20:00 228.1 F 0.0
25:00 224.7 F 0.0
30:00 221.9 F 0.0
Drying Complete

FILLING: 2:01
HEATING: 18:53
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:04
CYCLE COMPLETE

06/20/2023 12:11 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____