

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2758

Name: RITTER 1  
Sterilizer ID: V2178198  
01/10/2023 12:22 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 185.7 F 0.0  
2:00 177.8 F 0.0  
4:00 199.1 F 4.8  
6:00 217.5 F 9.9  
8:00 226.6 F 7.3  
10:00 237.2 F 11.2  
12:00 236.2 F 8.9  
14:00 239.1 F 9.5  
16:00 249.2 F 14.7  
18:00 261.9 F 21.7  
20:00 272.8 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.0  
1:00 272.5 F 28.7  
2:00 271.3 F 27.9  
3:00 272.4 F 28.8  
4:00 272.1 F 28.3  
  
Min 271.1 F 27.9  
Max 272.7 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.5 F 0.0  
5:00 228.8 F 0.0  
10:00 224.2 F 0.0  
15:00 208.8 F 0.0  
20:00 213.1 F 0.0  
25:00 217.6 F 0.0  
30:00 209.9 F 0.0  
Drying Complete

FILLING: 1:15  
HEATING: 20:21  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:58:50  
CYCLE COMPLETE

01/10/2023 01:20 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_