
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 006197
11:59 29/08/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:11
PRESSURIZING 3:11
STERILIZING 4:09
136.4 C 223kPa 4:09
136.2 C 223kPa 4:39
136.4 C 225kPa 5:09
136.4 C 224kPa 5:39
136.4 C 225kPa 6:09
136.2 C 221kPa 6:39
136.4 C 224kPa 7:09
136.2 C 220kPa 7:39
136.2 C 220kPa 7:39
VENTING 7:40
AIR DRYING 8:36
CYCLE COMPLETE 8:42

Digital Signature #
2EE6A18464130BF3
