
Midmark M11 - v1.0.5
Total Cycles: 3517

Name: RITTER2
Sterilizer ID: V2313228
04/11/2024 11:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.6 F 0.0
2:00 160.6 F 2.8
4:00 177.6 F 6.5
6:00 199.4 F 10.9
8:00 211.6 F 4.3
10:00 222.8 F 9.4
12:00 232.4 F 9.8
14:00 240.7 F 12.2
16:00 240.7 F 11.2
18:00 253.6 F 17.8
20:00 265.1 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.3 F 29.8
2:00 271.6 F 29.1
3:00 271.4 F 29.4
4:00 272.1 F 29.8

Min 271.0 F 28.9
Max 272.4 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.7 F 0.0
5:00 219.0 F 0.0
10:00 210.8 F 0.0
15:00 207.8 F 0.0
20:00 225.2 F 0.0
25:00 225.4 F 0.0
30:00 227.2 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 21:51
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:00:34
CYCLE COMPLETE

04/11/2024 12:32 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____