
Midmark M11 - v1.0.5
Total Cycles: 3942

Name: RITTER2
Sterilizer ID: V2313228
09/13/2024 09:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.9 F 0.0
2:00 71.6 F 0.0
4:00 77.4 F 0.0
6:00 125.2 F 0.0
8:00 147.2 F 0.2
10:00 169.6 F 0.9
12:00 195.9 F 4.3
14:00 219.7 F 8.0
16:00 230.8 F 11.8
18:00 227.0 F 6.5
20:00 237.0 F 11.1
22:00 233.0 F 7.9
24:00 244.5 F 12.2
26:00 244.7 F 12.7
28:00 256.6 F 19.0
30:00 267.2 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 272.2 F 28.5
1:00 272.5 F 29.3
2:00 271.5 F 28.4
3:00 272.7 F 29.2
4:00 271.4 F 28.5

Min 271.4 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.6 F 0.0
5:00 237.4 F 0.0
10:00 235.0 F 0.0
15:00 224.4 F 0.0
20:00 224.3 F 0.0
25:00 226.5 F 0.0
30:00 230.0 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 31:19
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:09:28
CYCLE COMPLETE

09/13/2024 10:12 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____