
Midmark M11 - v1.0.5
Total Cycles: 4387

Name: RITTER2
Sterilizer ID: V2313228
02/27/2025 01:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 105.7 F 0.0
2:00 105.5 F 0.7
4:00 114.2 F 2.4
6:00 138.6 F 5.2
8:00 161.6 F 8.8
10:00 192.5 F 12.8
12:00 206.1 F 5.3
14:00 219.6 F 9.2
16:00 231.7 F 13.5
18:00 236.1 F 10.0
20:00 237.5 F 9.4
22:00 245.8 F 13.8
24:00 256.7 F 19.8
26:00 267.3 F 26.3
28:00 272.1 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.8
1:00 271.8 F 29.7
2:00 271.8 F 29.7
3:00 271.7 F 29.7
4:00 272.4 F 29.9

Min 271.1 F 29.7
Max 273.1 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.9 F 0.0
5:00 226.1 F 0.0
10:00 217.0 F 0.0
15:00 215.8 F 0.0
20:00 177.8 F 0.0
25:00 168.0 F 0.0
30:00 167.2 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 28:06
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:06:48
CYCLE COMPLETE

02/27/2025 02:25 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____