
Midmark M11 - v1.0.0
Total Cycles: 778

Name: RITTER 1
Sterilizer ID: V2178198
04/08/2021 06:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.6 F 0.0
2:00 84.1 F 0.0
4:00 87.2 F 0.0
6:00 114.4 F 0.0
8:00 145.2 F 0.1
10:00 176.8 F 0.3
12:00 195.9 F 2.4
14:00 212.7 F 5.2
16:00 223.2 F 8.1
18:00 230.7 F 11.5
20:00 228.0 F 6.5
22:00 238.2 F 11.8
24:00 237.4 F 9.1
26:00 240.4 F 10.1
28:00 247.0 F 13.4
30:00 258.0 F 19.3
32:00 268.1 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 27.9
1:00 272.5 F 28.5
2:00 271.1 F 28.0
3:00 272.1 F 28.5
4:00 271.0 F 27.9

Min 271.0 F 27.8
Max 272.6 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 216.1 F 0.0
5:00 225.5 F 0.1
10:00 221.8 F 0.0
15:00 214.0 F 0.0
20:00 212.9 F 0.0
25:00 208.8 F 0.0
30:00 205.7 F 0.0

Drying Complete

FILLING: 0:54
HEATING: 33:19
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:11:29
CYCLE COMPLETE

04/08/2021 07:57 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____