
Midmark M11 - v1.0.0
Total Cycles: 2188

Name: RITTER 1
Sterilizer ID: V2178198
07/27/2022 02:54 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.4 F 0.0
2:00 179.6 F 4.1
4:00 192.6 F 8.8
6:00 208.3 F 9.6
8:00 227.6 F 8.4
10:00 239.5 F 13.5
12:00 237.8 F 10.4
14:00 235.0 F 7.9
16:00 248.8 F 14.9
18:00 260.7 F 21.5
20:00 271.5 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.6 F 29.2
2:00 271.7 F 28.5
3:00 272.1 F 29.2
4:00 272.3 F 29.0

Min 271.3 F 28.4
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.5 F 0.0
5:00 231.1 F 0.0
10:00 236.4 F 0.0
15:00 234.3 F 0.0
20:00 228.4 F 0.0
25:00 228.4 F 0.0
30:00 228.4 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 20:39
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:07
CYCLE COMPLETE

07/27/2022 03:53 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____