

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1280

Name: RITTER2  
Sterilizer ID: V2313228  
07/11/2022 10:06 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 180.5 F 0.0  
2:00 179.1 F 3.3  
4:00 185.0 F 7.2  
6:00 204.7 F 12.2  
8:00 218.2 F 6.6  
10:00 231.2 F 11.5  
12:00 232.5 F 8.3  
14:00 242.0 F 11.8  
16:00 244.0 F 12.7  
18:00 256.1 F 19.0  
20:00 266.6 F 25.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.4  
1:00 272.2 F 29.5  
2:00 272.2 F 29.9  
3:00 271.5 F 29.6  
4:00 272.2 F 29.5  
  
Min 271.5 F 29.2  
Max 272.4 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.4 F 0.0  
5:00 208.2 F 0.0  
10:00 223.1 F 0.0  
15:00 223.2 F 0.0  
20:00 222.8 F 0.0  
25:00 221.6 F 0.0  
30:00 219.1 F 0.0  
Drying Complete

FILLING: 1:05  
HEATING: 21:36  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:59:49  
CYCLE COMPLETE

07/11/2022 11:05 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_