

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2127

Name: RITTER2  
Sterilizer ID: V2313228  
03/08/2023 01:21 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 195.1 F 0.0  
2:00 195.6 F 4.2  
4:00 211.3 F 9.5  
6:00 225.6 F 7.3  
8:00 235.4 F 11.2  
10:00 234.7 F 8.9  
12:00 239.3 F 9.7  
14:00 248.8 F 15.0  
16:00 261.7 F 22.2  
18:00 272.3 F 29.8

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 29.0  
1:00 271.7 F 29.3  
2:00 272.9 F 29.9  
3:00 272.8 F 29.5  
4:00 272.1 F 29.1  
  
Min 271.3 F 28.8  
Max 273.1 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.4 F 0.0  
5:00 231.4 F 0.0  
10:00 222.0 F 0.0  
15:00 221.8 F 0.0  
20:00 223.2 F 0.0  
25:00 225.5 F 0.0  
30:00 220.7 F 0.0  
Drying Complete

FILLING: 1:60  
HEATING: 18:24  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:57:35  
CYCLE COMPLETE

03/08/2023 02:18 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_