
Midmark M11 - v1.0.5
Total Cycles: 3262

Name: RITTER2
Sterilizer ID: V2313228
12/28/2023 03:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 117.5 F 0.0
2:00 118.1 F 0.6
4:00 140.2 F 3.0
6:00 167.0 F 6.8
8:00 193.5 F 11.4
10:00 210.9 F 4.7
12:00 223.5 F 9.2
14:00 231.6 F 10.9
16:00 237.4 F 10.9
18:00 235.7 F 9.1
20:00 249.7 F 15.7
22:00 261.3 F 22.3
24:00 272.0 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.7 F 29.7
2:00 271.6 F 29.2
3:00 272.6 F 29.7
4:00 271.6 F 29.0

Min 271.5 F 29.0
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.0 F 0.0
5:00 228.5 F 0.0
10:00 237.8 F 0.0
15:00 234.4 F 0.0
20:00 223.3 F 0.0
25:00 223.7 F 0.0
30:00 223.4 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 24:33
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:02:46
CYCLE COMPLETE

12/28/2023 04:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____