
Midmark M11 - v1.0.0
Total Cycles: 4704

Name: RITTER 1
Sterilizer ID: V2178198
07/08/2024 10:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.8 F 0.0
2:00 158.4 F 2.2
4:00 160.4 F 5.4
6:00 187.5 F 9.2
8:00 205.1 F 6.0
10:00 217.7 F 6.5
12:00 230.1 F 11.2
14:00 231.7 F 6.7
16:00 241.7 F 12.5
18:00 239.9 F 10.5
20:00 252.5 F 16.8
22:00 263.7 F 23.2
24:00 272.0 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.4
1:00 272.3 F 29.1
2:00 272.0 F 28.6
3:00 272.0 F 28.9
4:00 272.4 F 28.9

Min 271.5 F 28.3
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.1 F 0.0
5:00 217.1 F 0.0
10:00 218.9 F 0.0
15:00 221.9 F 0.0
20:00 233.7 F 0.0
25:00 228.2 F 0.0
30:00 224.2 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 24:01
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:02:22
CYCLE COMPLETE

07/08/2024 11:30 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____