

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5230

Name: RITTER 1  
Sterilizer ID: V2178198  
01/09/2025 10:59 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 177.2 F 0.0  
2:00 174.2 F 2.6  
4:00 178.3 F 6.0  
6:00 197.8 F 10.5  
8:00 217.9 F 4.8  
10:00 226.7 F 8.2  
12:00 237.9 F 13.0  
14:00 236.8 F 10.0  
16:00 236.6 F 8.7  
18:00 248.0 F 14.3  
20:00 259.6 F 20.6  
22:00 269.8 F 27.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.3  
1:00 271.8 F 28.3  
2:00 272.0 F 28.5  
3:00 272.2 F 28.7  
4:00 272.3 F 29.0

Min 271.3 F 28.2  
Max 272.4 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.4 F 0.0  
5:00 220.2 F 0.0  
10:00 206.2 F 0.0  
15:00 210.4 F 0.0  
20:00 236.4 F 0.0  
25:00 234.4 F 0.0  
30:00 228.1 F 0.0  
Drying Complete

FILLING: 1:18  
HEATING: 22:54  
STERILIZING: 4:00  
VENTING: 3:48  
DRYING: 30:00  
TOTAL CYCLE: 01:02:00  
CYCLE COMPLETE

01/09/2025 12:00 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_