
Midmark M11 - v1.0.5
Total Cycles: 3218

Name: RITTER2
Sterilizer ID: V2313228
12/14/2023 03:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.0 F 0.0
2:00 168.3 F 3.3
4:00 188.9 F 7.7
6:00 208.7 F 13.1
8:00 221.9 F 7.8
10:00 235.9 F 13.3
12:00 240.2 F 11.0
14:00 238.7 F 9.9
16:00 253.0 F 17.0
18:00 265.1 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.6
1:00 272.7 F 29.5
2:00 271.9 F 28.8
3:00 272.0 F 29.5
4:00 272.4 F 29.1

Min 270.9 F 28.4
Max 273.0 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.4 F 0.0
5:00 216.3 F 0.0
10:00 200.3 F 0.0
15:00 199.4 F 0.0
20:00 196.1 F 0.0
25:00 195.3 F 0.0
30:00 192.9 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 19:50
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:58:29
CYCLE COMPLETE

12/14/2023 04:35 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____