
Midmark M11 - v1.0.5
Total Cycles: 1325

Name: RITTER2
Sterilizer ID: V2313228
07/20/2022 01:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.0 F 0.0
2:00 168.4 F 0.1
4:00 189.7 F 2.6
6:00 210.9 F 7.8
8:00 226.9 F 13.5
10:00 234.6 F 10.3
12:00 233.9 F 8.2
14:00 239.6 F 10.3
16:00 248.0 F 14.7
18:00 260.9 F 21.9
20:00 272.4 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.7 F 29.6
2:00 271.7 F 28.8
3:00 272.6 F 29.8
4:00 272.4 F 29.4

Min 271.5 F 28.8
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.4 F 0.0
5:00 231.9 F 0.0
10:00 239.1 F 0.0
15:00 235.3 F 0.0
20:00 228.3 F 0.0
25:00 224.0 F 0.0
30:00 226.8 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 20:30
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 00:58:42
CYCLE COMPLETE

07/20/2022 02:27 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____