
Midmark M11 - v1.0.0
Total Cycles: 2499

Name: RITTER 1
Sterilizer ID: V2178198
11/02/2022 09:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.1 F 0.0
2:00 165.5 F 2.3
4:00 175.2 F 5.9
6:00 188.5 F 10.4
8:00 210.6 F 4.0
10:00 223.7 F 7.9
12:00 237.2 F 12.8
14:00 236.5 F 9.4
16:00 239.5 F 9.2
18:00 247.2 F 13.7
20:00 258.8 F 20.1
22:00 269.7 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.5 F 28.9
2:00 271.5 F 28.1
3:00 272.6 F 29.0
4:00 272.0 F 28.5

Min 271.3 F 28.1
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.3 F 0.0
5:00 228.3 F 0.0
10:00 225.4 F 0.0
15:00 246.0 F 0.0
20:00 234.1 F 0.0
25:00 228.3 F 0.0
30:00 227.7 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 22:58
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:01:01
CYCLE COMPLETE

11/02/2022 10:23 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____