
Midmark M11 - v1.0.0
Total Cycles: 4429

Name: RITTER 1
Sterilizer ID: V2178198
04/05/2024 09:31 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.9 F 0.0
2:00 165.2 F 1.5
4:00 168.7 F 4.7
6:00 187.6 F 8.3
8:00 206.1 F 12.9
10:00 224.0 F 6.4
12:00 232.1 F 10.7
14:00 231.8 F 7.3
16:00 242.1 F 12.6
18:00 240.3 F 10.7
20:00 252.7 F 16.7
22:00 263.3 F 23.0
24:00 272.3 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.4 F 28.9
2:00 272.5 F 29.4
3:00 271.4 F 28.7
4:00 272.7 F 29.2

Min 271.2 F 28.6
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.1 F 0.0
5:00 220.7 F 0.0
10:00 227.6 F 0.0
15:00 225.6 F 0.0
20:00 222.9 F 0.0
25:00 225.7 F 0.0
30:00 225.2 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 24:55
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:03:52
CYCLE COMPLETE

04/05/2024 10:34 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____