

-----  
Midmark M11 - v1.0.0  
Total Cycles: 855

Name: RITTER 1  
Sterilizer ID: V2178198  
05/31/2021 02:45 PM

BEGIN PACKS CYCLE  
Temp: 250 Degrees F  
Time: 30 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 184.4 F 0.0  
2:00 183.0 F 3.5  
4:00 213.7 F 4.1  
6:00 222.8 F 5.0  
8:00 225.4 F 4.1  
10:00 229.5 F 5.5  
12:00 237.8 F 9.0  
14:00 250.7 F 15.2

STERILIZING  
mm:ss Degrees PSI  
0:00 251.9 F 15.6  
1:00 252.4 F 16.1  
2:00 251.6 F 15.5  
3:00 252.4 F 16.2  
4:00 251.9 F 15.7  
5:00 251.9 F 16.0  
6:00 252.4 F 16.1  
7:00 251.8 F 15.7  
8:00 251.5 F 15.7  
9:00 252.5 F 16.2  
10:00 252.3 F 16.0  
11:00 251.9 F 15.8  
12:00 251.6 F 15.6  
13:00 251.8 F 16.0  
14:00 252.5 F 16.2  
15:00 252.3 F 16.1  
16:00 251.9 F 15.9  
17:00 251.8 F 15.7  
18:00 251.5 F 15.6  
19:00 251.8 F 15.9  
20:00 252.5 F 16.3  
21:00 252.6 F 16.2  
22:00 252.2 F 16.0  
23:00 251.9 F 15.9  
24:00 251.6 F 15.7  
25:00 251.5 F 15.6  
26:00 251.6 F 15.8  
27:00 252.4 F 16.3  
28:00 252.7 F 16.2  
29:00 252.5 F 16.1  
30:00 252.1 F 16.0

Min 251.3 F 15.5  
Max 252.8 F 16.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.5 F 0.0  
5:00 196.7 F 0.0  
10:00 166.1 F 0.0  
15:00 170.9 F 0.0  
20:00 151.3 F 0.0  
Dry Cycle Aborted

FILLING: 2:36  
HEATING: 14:37  
STERILIZING: 30:00  
VENTING: 2:55  
DRYING: 20:38

TOTAL CYCLE: 01:10:46

05/31/2021 03:55 PM

ACCEPTED: X

REJECTED:

Operator: U01