
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001114
16:18 08/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:16
PRESSURIZING 4:18
STERILIZING 5:53
136.4 C 222kPa 5:53
136.2 C 224kPa 6:23
136.4 C 224kPa 6:53
136.4 C 225kPa 7:23
136.4 C 225kPa 7:53
136.3 C 223kPa 8:23
136.3 C 225kPa 8:53
136.4 C 225kPa 9:23
136.4 C 225kPa 9:23
VENTING 9:23
AIR DRYING 10:25
CYCLE COMPLETE 25:26

Digital Signature #
A9ECD121104B21FE
