

-----  
Midmark M11 - v1.0.5  
Total Cycles: 151

Name: RITTER2  
Sterilizer ID: V2313228  
06/28/2021 02:39 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 153.4 F 0.0  
2:00 152.4 F 1.8  
4:00 155.6 F 4.8  
6:00 177.3 F 8.7  
8:00 197.7 F 12.9  
10:00 210.6 F 7.0  
12:00 223.9 F 11.6  
14:00 230.3 F 8.7  
16:00 238.8 F 11.4  
18:00 242.9 F 13.1  
20:00 254.9 F 19.1  
22:00 264.8 F 25.4  
24:00 271.8 F 30.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.9  
1:00 271.5 F 30.2  
2:00 272.0 F 30.5  
3:00 272.3 F 30.8  
4:00 272.8 F 30.9

Min 271.2 F 29.7  
Max 273.6 F 30.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.0 F 0.0  
5:00 215.5 F 0.0  
10:00 224.4 F 0.0  
15:00 216.4 F 0.0  
20:00 211.4 F 0.0  
25:00 211.4 F 0.0  
30:00 206.0 F 0.0  
Drying Complete

FILLING: 1:30  
HEATING: 24:11  
STERILIZING: 4:00  
VENTING: 3:31  
DRYING: 30:00  
TOTAL CYCLE: 01:03:12  
CYCLE COMPLETE

06/28/2021 03:48 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01