
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000290
10:01 16/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:59
PRESSURIZING 3:04
STERILIZING 4:09
136.0 C 228kPa 4:09
136.4 C 233kPa 4:39
136.4 C 233kPa 5:09
136.2 C 230kPa 5:39
136.1 C 230kPa 6:09
136.1 C 228kPa 6:39
136.4 C 233kPa 7:09
136.2 C 231kPa 7:39
136.2 C 231kPa 7:39
VENTING 7:39
AIR DRYING 8:30
CYCLE COMPLETE 20:54

Digital Signature #
BE3F54EA4994DC03
