
Midmark M11 - v1.0.0
Total Cycles: 4917

Name: RITTER 1
Sterilizer ID: V2178198
09/18/2024 03:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 109.9 F 0.0
2:00 112.4 F 0.0
4:00 133.0 F 0.1
6:00 166.4 F 0.4
8:00 193.0 F 3.2
10:00 214.5 F 7.2
12:00 226.0 F 11.4
14:00 228.1 F 6.4
16:00 240.8 F 11.8
18:00 237.0 F 9.2
20:00 238.6 F 9.4
22:00 248.4 F 14.2
24:00 260.3 F 20.8
26:00 271.2 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.0
1:00 272.5 F 29.0
2:00 271.6 F 28.4
3:00 272.5 F 28.7
4:00 272.2 F 28.9

Min 271.0 F 27.9
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.6 F 0.0
5:00 226.5 F 0.0
10:00 221.6 F 0.0
15:00 219.0 F 0.0
20:00 227.7 F 0.0
25:00 231.9 F 0.0
30:00 222.8 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 26:33
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:05:01
CYCLE COMPLETE

09/18/2024 04:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____