
Midmark M11 - v1.0.5
Total Cycles: 2651

Name: RITTER2
Sterilizer ID: V2313228
07/17/2023 08:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.8 F 0.0
2:00 78.0 F 0.0
4:00 87.9 F 0.0
6:00 105.1 F 0.0
8:00 153.8 F 0.3
10:00 180.8 F 3.3
12:00 190.6 F 6.9
14:00 202.6 F 10.4
16:00 218.7 F 7.9
18:00 224.1 F 7.9
20:00 233.5 F 12.0
22:00 233.2 F 8.6
24:00 241.7 F 11.7
26:00 243.9 F 12.7
28:00 255.7 F 18.8
30:00 266.1 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.6 F 29.8
2:00 272.0 F 29.1
3:00 271.4 F 29.2
4:00 272.7 F 29.8

Min 271.2 F 28.8
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.4 F 0.0
5:00 211.8 F 0.0
10:00 209.6 F 0.0
15:00 202.5 F 0.0
20:00 207.8 F 0.0
25:00 210.3 F 0.0
30:00 205.6 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 31:43
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:09:42
CYCLE COMPLETE

07/17/2023 09:59 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____