
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 004979
12:25 23/01/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:53
PRESSURIZING 1:57
STERILIZING 3:06
136.4 C 220kPa 3:06
136.4 C 225kPa 3:36
136.3 C 223kPa 4:06
136.4 C 225kPa 4:36
136.4 C 225kPa 5:06
136.4 C 219kPa 5:36
136.5 C 225kPa 6:06
136.4 C 223kPa 6:36
136.4 C 223kPa 6:36
VENTING 6:36
AIR DRYING 7:27
CYCLE COMPLETE 7:40

Digital Signature #
4E7C253E25C0E6CA
