

-----  
Midmark M11 - v1.0.5  
Total Cycles: 895

Name: RITTER2  
Sterilizer ID: V2313228  
03/23/2022 09:41 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 105.1 F 0.0  
2:00 106.2 F 0.0  
4:00 116.1 F 0.0  
6:00 159.6 F 0.3  
8:00 197.7 F 1.0  
10:00 211.1 F 5.2  
12:00 224.3 F 9.3  
14:00 235.2 F 11.1  
16:00 238.5 F 11.1  
18:00 237.1 F 9.4  
20:00 239.0 F 9.4  
22:00 250.6 F 15.6  
24:00 262.9 F 22.8  
26:00 272.3 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.5  
1:00 272.6 F 29.4  
2:00 272.0 F 28.6  
3:00 271.5 F 28.9  
4:00 272.6 F 29.3

Min 271.2 F 28.4  
Max 272.7 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.0 F 0.0  
5:00 218.0 F 0.0  
10:00 217.5 F 0.0  
15:00 209.1 F 0.0  
20:00 215.8 F 0.0  
25:00 220.7 F 0.0  
30:00 225.9 F 0.0  
Drying Complete

FILLING: 0:60  
HEATING: 26:12  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 30:00  
TOTAL CYCLE: 01:04:18  
CYCLE COMPLETE

03/23/2022 10:45 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_