
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002636
8:48 31/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:07
PRESSURIZING 2:13
STERILIZING 3:14
136.0 C 220kPa 3:14
136.4 C 226kPa 3:44
136.4 C 222kPa 4:14
136.1 C 223kPa 4:44
136.4 C 226kPa 5:14
136.4 C 225kPa 5:44
136.4 C 224kPa 6:14
136.0 C 222kPa 6:44
136.0 C 222kPa 6:44
VENTING 6:44
AIR DRYING 7:36
CYCLE COMPLETE 12:26

Digital Signature #
88F0D0FB8E2F940E
