
Midmark M11 - v1.0.5
Total Cycles: 3535

Name: RITTER2
Sterilizer ID: V2313228
04/17/2024 09:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 136.1 F 0.0
2:00 136.7 F 1.2
4:00 137.6 F 3.7
6:00 157.5 F 7.0
8:00 181.1 F 10.8
10:00 208.3 F 4.6
12:00 218.4 F 7.4
14:00 230.1 F 11.7
16:00 229.8 F 8.2
18:00 241.7 F 13.5
20:00 241.7 F 12.0
22:00 253.7 F 18.0
24:00 264.5 F 24.4
26:00 271.2 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.2
1:00 272.6 F 29.8
2:00 271.5 F 29.1
3:00 271.8 F 29.8
4:00 272.2 F 29.6

Min 271.1 F 29.0
Max 272.6 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.3 F 0.0
5:00 211.7 F 0.0
10:00 215.2 F 0.0
15:00 214.0 F 0.0
20:00 216.9 F 0.0
25:00 217.3 F 0.0
30:00 220.1 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 26:03
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:04:36
CYCLE COMPLETE

04/17/2024 10:56 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____