

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 000630  
7:36 15/10/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:32  
PRESSURIZING 2:36  
STERILIZING 3:42  
136.4 C 222kPa 3:42  
136.4 C 224kPa 4:12  
136.1 C 222kPa 4:42  
136.1 C 221kPa 5:12  
136.4 C 220kPa 5:42  
136.2 C 223kPa 6:12  
136.4 C 222kPa 6:42  
136.2 C 222kPa 7:12  
136.2 C 222kPa 7:12  
VENTING 7:12  
AIR DRYING 8:05  
CYCLE COMPLETE 8:21

-----  
Digital Signature #  
83A07BC243B14E1E  
-----