

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.6uS / 0.3ppm  
CYCLE NUMBER 000505  
14:53 07/05/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:12  
PRESSURIZING 4:13  
STERILIZING 5:32  
136.4 C 232kPa 5:32  
136.4 C 233kPa 6:02  
136.3 C 232kPa 6:32  
136.3 C 232kPa 7:02  
136.3 C 233kPa 7:32  
136.4 C 231kPa 8:02  
136.3 C 232kPa 8:32  
136.4 C 233kPa 9:02  
136.4 C 233kPa 9:02  
VENTING 9:02  
AIR DRYING 10:00  
CYCLE COMPLETE 10:23  
-----

Digital Signature #  
AAECA91CE55C99B9  
-----