
Midmark M11 - v1.0.0
Total Cycles: 1732

Name: RITTER 1
Sterilizer ID: V2178198
03/23/2022 07:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.5 F 0.0
2:00 73.4 F 0.0
4:00 81.3 F 0.0
6:00 113.2 F 0.3
8:00 147.7 F 2.1
10:00 172.4 F 4.6
12:00 191.3 F 8.2
14:00 207.3 F 12.2
16:00 216.9 F 5.0
18:00 226.9 F 9.2
20:00 238.2 F 13.0
22:00 237.3 F 10.1
24:00 235.8 F 8.3
26:00 247.7 F 14.0
28:00 259.5 F 20.5
30:00 270.1 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 271.9 F 28.2
2:00 272.3 F 28.6
3:00 272.6 F 28.9
4:00 271.9 F 28.6

Min 271.4 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.6 F 0.0
5:00 218.0 F 0.0
10:00 219.3 F 0.0
15:00 218.6 F 0.0
20:00 225.0 F 0.0
25:00 230.4 F 0.0
30:00 229.1 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 30:51
STERILIZING: 4:00
VENTING: 2:58
DRYING: 30:00
TOTAL CYCLE: 01:08:42
CYCLE COMPLETE

03/23/2022 08:28 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____