

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 001427  
7:59 31/03/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:09  
PRESSURIZING 3:12  
STERILIZING 4:16  
136.4 C 220kPa 4:16  
136.4 C 225kPa 4:46  
136.4 C 222kPa 5:16  
136.3 C 223kPa 5:46  
136.4 C 224kPa 6:16  
136.2 C 219kPa 6:46  
136.4 C 221kPa 7:16  
136.4 C 223kPa 7:46  
136.4 C 223kPa 7:46  
VENTING 7:46  
AIR DRYING 8:36  
CYCLE COMPLETE 8:42

-----  
Digital Signature #  
50B06D6CD8CC329D  
-----