
Midmark M11 - v1.0.5
Total Cycles: 1071

Name: RITTER2
Sterilizer ID: V2313228
05/04/2022 11:22 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.6 F 0.0
2:00 181.3 F 3.7
4:00 190.4 F 8.3
6:00 210.5 F 11.7
8:00 224.0 F 8.5
10:00 235.4 F 11.6
12:00 239.2 F 11.7
14:00 238.9 F 10.5
16:00 252.2 F 17.2
18:00 263.8 F 24.0
20:00 271.8 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.3
1:00 272.6 F 30.0
2:00 271.8 F 29.3
3:00 271.7 F 29.8
4:00 272.7 F 29.9

Min 271.3 F 29.2
Max 272.7 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.6 F 0.0
5:00 214.4 F 0.0
10:00 223.1 F 0.0
15:00 221.1 F 0.0
20:00 218.2 F 0.0
25:00 216.1 F 0.0
30:00 214.7 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 20:08
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:53
CYCLE COMPLETE

05/04/2022 12:20 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____