

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1238

Name: RITTER2  
Sterilizer ID: V2313228  
06/29/2022 07:18 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 74.3 F 0.0  
2:00 77.3 F 0.0  
4:00 104.1 F 0.0  
6:00 113.2 F 0.1  
8:00 150.6 F 0.4  
10:00 175.6 F 2.6  
12:00 200.3 F 5.8  
14:00 207.8 F 9.7  
16:00 220.6 F 11.6  
18:00 227.4 F 8.3  
20:00 239.1 F 12.9  
22:00 237.6 F 10.2  
24:00 237.9 F 8.7  
26:00 248.4 F 14.9  
28:00 259.5 F 21.1  
30:00 269.5 F 27.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.9  
1:00 272.7 F 29.8  
2:00 272.1 F 29.1  
3:00 271.3 F 29.0  
4:00 273.0 F 29.8

Min 271.2 F 28.7  
Max 273.0 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.9 F 0.0  
5:00 196.5 F 0.0  
10:00 216.6 F 0.0  
15:00 237.0 F 0.0  
20:00 227.5 F 0.0  
25:00 227.9 F 0.0  
30:00 228.9 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 31:05  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:09:01  
CYCLE COMPLETE

06/29/2022 08:26 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_