
Midmark M11 - v1.0.5
Total Cycles: 704

Name: RITTER2
Sterilizer ID: V2313228
01/19/2022 10:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.6 F 0.0
2:00 145.3 F 1.9
4:00 155.5 F 5.0
6:00 180.4 F 9.3
8:00 203.8 F 7.0
10:00 218.5 F 8.3
12:00 232.2 F 12.1
14:00 238.1 F 11.4
16:00 238.2 F 10.1
18:00 251.7 F 16.9
20:00 263.5 F 23.9
22:00 271.7 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.4
1:00 272.5 F 30.1
2:00 271.5 F 29.3
3:00 271.7 F 30.0
4:00 272.4 F 29.9

Min 271.2 F 29.2
Max 272.7 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.2 F 0.0
5:00 210.9 F 0.0
10:00 224.2 F 0.0
15:00 219.9 F 0.0
20:00 216.2 F 0.0
25:00 211.9 F 0.0
30:00 216.7 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 22:08
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:00:20
CYCLE COMPLETE

01/19/2022 11:33 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____