

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5036

Name: RITTER 1  
Sterilizer ID: V2178198  
10/30/2024 08:28 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.2 F 0.0  
2:00 92.8 F 0.0  
4:00 95.5 F 0.0  
6:00 115.0 F 0.1  
8:00 144.4 F 0.3  
10:00 177.7 F 0.8  
12:00 197.5 F 4.2  
14:00 219.6 F 7.4  
16:00 226.1 F 11.0  
18:00 226.2 F 5.9  
20:00 235.4 F 9.7  
22:00 238.2 F 9.8  
24:00 243.7 F 11.9  
26:00 240.3 F 10.2  
28:00 253.3 F 16.6  
30:00 264.4 F 23.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 27.9  
1:00 272.5 F 28.5  
2:00 271.4 F 27.8  
3:00 272.6 F 28.6  
4:00 271.8 F 27.9

Min 271.3 F 27.7  
Max 272.7 F 28.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 177.0 F 0.0  
5:00 217.3 F 0.0  
10:00 212.1 F 0.0  
15:00 242.3 F 0.0  
20:00 224.5 F 0.0  
25:00 228.9 F 0.0  
30:00 227.6 F 0.0  
Drying Complete

FILLING: 0:41  
HEATING: 31:60  
STERILIZING: 4:00  
VENTING: 3:25  
DRYING: 30:00  
TOTAL CYCLE: 01:10:06  
CYCLE COMPLETE

10/30/2024 09:38 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_