

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1917

Name: RITTER2  
Sterilizer ID: V2313228  
01/11/2023 09:49 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 131.2 F 0.0  
2:00 133.3 F 1.4  
4:00 143.4 F 4.2  
6:00 170.3 F 8.2  
8:00 196.3 F 13.1  
10:00 211.8 F 6.4  
12:00 225.6 F 11.4  
14:00 231.2 F 8.1  
16:00 241.1 F 11.6  
18:00 244.7 F 13.1  
20:00 257.1 F 19.7  
22:00 268.0 F 26.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.2  
1:00 272.1 F 29.4  
2:00 271.8 F 29.7  
3:00 271.8 F 29.3  
4:00 271.9 F 29.8

Min 271.3 F 29.0  
Max 272.3 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.6 F 0.0  
5:00 212.2 F 0.0  
10:00 224.4 F 0.0  
15:00 231.3 F 0.0  
20:00 231.9 F 0.0  
25:00 226.5 F 0.0  
30:00 225.2 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 23:14  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 01:01:17  
CYCLE COMPLETE

01/11/2023 10:50 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_