
Midmark M11 - v1.0.5
Total Cycles: 2238

Name: RITTER2
Sterilizer ID: V2313228
04/12/2023 09:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.6 F 0.0
2:00 183.2 F 3.0
4:00 195.0 F 8.1
6:00 213.7 F 10.0
8:00 227.8 F 9.9
10:00 232.7 F 8.3
12:00 240.8 F 11.1
14:00 246.5 F 14.2
16:00 259.6 F 21.4
18:00 271.3 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.2
1:00 271.8 F 29.7
2:00 272.5 F 29.8
3:00 271.7 F 29.2
4:00 271.3 F 29.4

Min 271.2 F 29.0
Max 272.7 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.7 F 0.0
5:00 228.0 F 0.0
10:00 229.6 F 0.0
15:00 234.6 F 0.0
20:00 229.9 F 0.0
25:00 229.6 F 0.0
30:00 225.7 F 0.0
Drying Complete

FILLING: 1:58
HEATING: 18:35
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:57:48
CYCLE COMPLETE

04/12/2023 10:49 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____