
Midmark M11 - v1.0.5
Total Cycles: 654

Name: RITTER2
Sterilizer ID: V2313228
12/20/2021 01:51 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.4 F 0.0
2:00 187.1 F 4.3
4:00 204.2 F 10.1
6:00 220.1 F 6.2
8:00 236.0 F 12.1
10:00 237.2 F 10.1
12:00 235.7 F 8.7
14:00 251.0 F 16.1
16:00 263.3 F 23.3
18:00 271.5 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.7
1:00 272.7 F 29.5
2:00 271.3 F 28.6
3:00 272.5 F 29.7
4:00 271.9 F 29.0

Min 271.1 F 28.6
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.8 F 0.0
5:00 230.7 F 0.0
10:00 220.0 F 0.0
15:00 217.4 F 0.0
20:00 204.5 F 0.0
25:00 193.4 F 0.0
30:00 198.0 F 0.0
Drying Complete

FILLING: 2:15
HEATING: 18:06
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 00:57:23
CYCLE COMPLETE

12/20/2021 02:55 PM
ACCEPTED: X
REJECTED:
Operator: U02