
Midmark M11 - v1.0.0
Total Cycles: 3372

Name: RITTER 1
Sterilizer ID: V2178198
06/12/2023 10:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 210.1 F 0.0
2:00 196.8 F 3.9
4:00 200.4 F 8.6
6:00 220.9 F 10.2
8:00 228.4 F 9.4
10:00 236.4 F 8.9
12:00 241.6 F 12.3
14:00 240.3 F 10.7
16:00 253.3 F 17.2
18:00 264.6 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.5 F 29.1
2:00 271.4 F 28.6
3:00 272.5 F 29.1
4:00 271.4 F 28.5

Min 271.3 F 28.4
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.0 F 0.0
5:00 225.8 F 0.0
10:00 238.3 F 0.0
15:00 242.9 F 0.0
20:00 230.7 F 0.0
25:00 227.9 F 0.0
30:00 215.1 F 0.0
Drying Complete

FILLING: 1:54
HEATING: 19:54
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:02
CYCLE COMPLETE

06/12/2023 11:15 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____