

-----  
Midmark M11 - v1.0.5  
Total Cycles: 966

Name: RITTER2  
Sterilizer ID: V2313228  
04/07/2022 08:15 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 187.9 F 0.0  
2:00 181.4 F 0.5  
4:00 185.1 F 3.7  
6:00 205.8 F 9.1  
8:00 226.0 F 8.8  
10:00 236.3 F 10.9  
12:00 234.8 F 8.9  
14:00 238.9 F 10.0  
16:00 249.1 F 15.2  
18:00 261.6 F 22.3  
20:00 272.8 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.2  
1:00 272.7 F 29.5  
2:00 272.4 F 29.9  
3:00 272.4 F 29.2  
4:00 272.7 F 29.9  
  
Min 271.5 F 29.1  
Max 273.0 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.2 F 0.0  
5:00 222.0 F 0.0  
10:00 215.1 F 0.0  
15:00 225.2 F 0.0  
20:00 212.7 F 0.0  
25:00 217.6 F 0.0  
30:00 215.3 F 0.0  
Drying Complete

FILLING: 1:30  
HEATING: 21:04  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:59:45  
CYCLE COMPLETE

04/07/2022 09:19 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02