

-----  
Midmark M11 - v1.0.5  
Total Cycles: 594

Name: RITTER2  
Sterilizer ID: V2313228  
12/01/2021 11:08 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 122.7 F 0.0  
2:00 125.4 F 1.5  
4:00 151.1 F 4.6  
6:00 180.6 F 8.7  
8:00 200.5 F 13.3  
10:00 216.4 F 7.3  
12:00 231.6 F 12.6  
14:00 236.0 F 9.6  
16:00 237.2 F 9.3  
18:00 247.9 F 14.4  
20:00 259.6 F 21.1  
22:00 270.2 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.8  
1:00 272.2 F 29.2  
2:00 272.5 F 29.7  
3:00 271.5 F 28.7  
4:00 272.4 F 29.5

Min 271.3 F 28.7  
Max 272.9 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.2 F 0.0  
5:00 226.8 F 0.0  
10:00 228.4 F 0.0  
15:00 216.9 F 0.0  
20:00 213.7 F 0.0  
25:00 214.6 F 0.0  
30:00 214.8 F 0.0  
Drying Complete

FILLING: 1:18  
HEATING: 22:47  
STERILIZING: 4:00  
VENTING: 3:01  
DRYING: 30:00  
TOTAL CYCLE: 01:01:06  
CYCLE COMPLETE

12/01/2021 12:36 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02