

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4101

Name: RITTER2  
Sterilizer ID: V2313228  
11/08/2024 10:17 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.1 F 0.0  
2:00 71.2 F 0.0  
4:00 77.4 F 0.0  
6:00 107.8 F 0.0  
8:00 148.9 F 0.3  
10:00 173.6 F 0.3  
12:00 191.4 F 1.7  
14:00 208.5 F 5.1  
16:00 220.0 F 8.4  
18:00 229.9 F 12.1  
20:00 227.7 F 7.9  
22:00 239.2 F 12.6  
24:00 238.0 F 10.2  
26:00 237.0 F 9.0  
28:00 248.3 F 14.7  
30:00 259.0 F 20.7  
32:00 269.4 F 27.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.9  
1:00 272.7 F 29.5  
2:00 271.5 F 28.9  
3:00 272.6 F 29.5  
4:00 271.6 F 28.8  
  
Min 271.3 F 28.7  
Max 272.7 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.9 F 0.0  
5:00 194.6 F 0.0  
10:00 201.2 F 0.0  
15:00 203.6 F 0.0  
20:00 204.4 F 0.0  
25:00 210.8 F 0.0  
30:00 215.5 F 0.0  
  
Drying Complete

FILLING: 0:50  
HEATING: 33:03  
STERILIZING: 4:00  
VENTING: 3:34  
DRYING: 30:00  
TOTAL CYCLE: 01:11:28  
CYCLE COMPLETE

11/08/2024 11:28 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_