
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000111
10:17 03/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:30
PRESSURIZING 4:30
STERILIZING 5:57
136.5 C 225kPa 5:57
136.4 C 227kPa 6:27
136.4 C 225kPa 6:57
136.4 C 226kPa 7:27
136.4 C 225kPa 7:57
136.4 C 226kPa 8:27
136.2 C 224kPa 8:57
136.2 C 224kPa 9:27
136.2 C 224kPa 9:27
VENTING 9:27
AIR DRYING 10:28
CYCLE COMPLETE 25:28

Digital Signature #
78E98406EB7CCE1C
