
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002000
13:09 12/07/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:31
PRESSURIZING 3:32
STERILIZING 4:46
136.4 C 220kPa 4:46
136.3 C 224kPa 5:16
136.4 C 225kPa 5:46
136.4 C 224kPa 6:16
136.3 C 224kPa 6:46
136.3 C 225kPa 7:16
136.1 C 223kPa 7:46
136.4 C 222kPa 8:16
136.4 C 222kPa 8:16
VENTING 8:16
AIR DRYING 9:12
CYCLE COMPLETE 9:14

Digital Signature #
A705B75623213B38
