

-----  
Midmark M11 - v1.0.5  
Total Cycles: 734

Name: RITTER2  
Sterilizer ID: V2313228  
01/31/2022 09:40 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.6 F 0.0  
2:00 85.6 F 0.0  
4:00 82.2 F 0.0  
6:00 109.3 F 0.1  
8:00 148.3 F 0.3  
10:00 178.2 F 3.0  
12:00 196.7 F 6.5  
14:00 215.7 F 10.4  
16:00 222.8 F 7.3  
18:00 230.5 F 10.0  
20:00 234.8 F 9.2  
22:00 243.2 F 12.9  
24:00 242.8 F 11.9  
26:00 255.5 F 18.5  
28:00 266.7 F 25.4  
30:00 273.0 F 29.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.2  
1:00 272.8 F 29.6  
2:00 272.1 F 29.5  
3:00 272.7 F 29.5  
4:00 272.0 F 29.4

Min 271.8 F 29.0  
Max 272.9 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 218.7 F 0.0  
5:00 230.5 F 0.0  
10:00 211.8 F 0.0  
15:00 199.0 F 0.0  
20:00 195.6 F 0.0  
25:00 206.8 F 0.0  
30:00 209.1 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 30:18  
STERILIZING: 4:00  
VENTING: 3:00  
DRYING: 30:00  
TOTAL CYCLE: 01:08:12  
CYCLE COMPLETE

01/31/2022 10:48 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_